

CERTIFICATE OF ANALYSIS
“Sterile Surgical Radiation Protection Gloves(Lead-Free): Type UltraThin”: Size – 8.5

Batch Number	20L20031880		
Article Code	X1D12125.	Date of testing:	22 April 2018
Order quantity	120 pairs	Invoice Number	3165002691
Date of Manufacturing	April 2018	Date of Expiry	February 2021
Property	U/M	Limits	Results
Protein Level (type tested)	µg/g	≤ 50	16
Residual Powder (type tested)	mg/glove	≤ 2	<2
Pin Holes (1000ml water leak test)	Pass / Fail	As per G-1, AQL 2.5	Pass
Tensile Strength	MPa	≥ 14	15.0
Ultimate Elongation	%	Min. 560	610
Dimensions	Pass / Fail	As per ASTM D3577/EN 420	Pass
Attenuation	%	Min 41 %	44.0 %

Verified by & Approved by



Head – Q & R

Kiran Medical Systems

A division of Trivitron Healthcare Private Limited

D-117, TTC Industrial Area, Village Shirvane, P.O. Nerul, Dist. Thane, Navi Mumbai 400 706, India
 Ph: +91 22 27630942 | Fax: +91 22 27671790 | Web: www.kiranxray.com

Regd. Office: Trivitron Healthcare Private Limited

Trivitron Sapthagiri Bhawan, No.15 IV Street, Abhiramapuram, Chennai 600 018, Tamil Nadu, India
 Ph: +91 - 44 - 2498 5050 (6 lines) | Fax: +91 - 44 - 2498 5757 | Web: www.trivitron.com | CIN: U85110TN1998PTC040515